Tennessee Immunization Program (TIP)

Temperature Log for Freezer - Celsius

DAYS 1 - 15

Monitor temperatures closely!

- 1. Write your initials below in "Staff Initials," and note the time the temperature was checked.
- 2. Record temps twice each work day.
- 3. Record the min/max temps once each work day preferably in the morning.
- 4. Put an "X" in the row that corresponds to the refrigerator's temperature.
- 5. If any out of range temps, see instructions to the right.
- 6. Keep all VFC documents for 3 years.

Month/Year	VFC PIN#
Facility Name	
Take action if temp	is out of range (above -15°C o
below -50°C)	

- 1. Label vaccine "do not use," and store it under proper conditions as quickly as possible.
- 2. Do not discard vaccines unless directed to by TIP and/or vaccine manufacturer(s).
- 3. Record the out of range temp and the room temp in the "Action" box (which correlates with the date) on the bottom of the log.
- 4. Notify TIP at 1-800-404-3006 immediately for guidance.

																4.	NOU	IY IIP	at I-	6UU-4	104-3	006 ir	шпе	Jiatei	y ioi	guiuc	irice.			
DayofMonth		1		2		3	4	4		5	(6		7	8	3	!	9	1	0	1	1	1	2	1	3	1	4	1	15
Staff Initials																														
Exact Time Read	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm
Min/MaxTemp (since previous reading)																														
Danger! Temperatu	res a	bove	-15ºC	are t	00 W	arm! ۱	Nrite	any o	ut-of	-range	e tem	ps an	d roc	om tei	mp oı	n the	lines	belov	w and	call '	TIP in	nmedi	ately	! 1-8	00-40	4-300	6			
ှ -15°C																														
-16°C																														
-17°C																														
-18°C																														
-19°C																														
-20°C																														
-21°C																														
-22°C																														
-50°C to -23°C																														
Out-of-range temps (above -15°C or below -50°C)																														
Room Temperature																														

Please complete all information for each excursion outside the proper temperature range; above -15°C or below -50°C.

Day of Month	Time	Temperature	Actions Taken	Signature

Tennessee Immunization Program (TIP)

Temperature Log for Freezer - Celsius

DAYS 16 - 31

Monitor temperatures closely!

- 1. Write your initials below in "Staff Initials," and note the time the temperature was checked.
- 2. Record temps twice each work day.
- 3. Record the min/max temps once each work day preferably in the morning.
- 4. Put an "X" in the row that corresponds to the refrigerator's temperature.
- 5. If any out of range temps, see instructions to the right.
- 6. Keep all VFC documents for 3 years.

Month/Year	VFC PIN#
Facility Name	
Take action if ten	np is out of range (above -15°C or
helow -50°C)	

- 1. Label vaccine "do not use," and store it under proper conditions as quickly as possible.
- 2. Do not discard vaccines unless directed to by TIP and/or vaccine manufacturer(s).
- 3. Record the out of range temp and the room temp in the "Action" box (which correlates with the date) on the bottom of the log.
- 4. Notify TIP at 1-800-404-3006 immediately for guidance.

																		4.	. NOU	I y I I F	' at 1-	000-	+04-3	000 111	iiiie	Jiatei	y ioi ,	guiud	ince.			
Da	yofMonth	1	16	•	17		18	1	9	2	20	2	21	2	22	2	23	2	.4	2	25	2	6	27	7	2	8	2	29	30		31
Sta	ff Initials																															
Exa	act Time Read	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am	pm	am pm	aı	m pm
	n/MaxTemp ce previous reading)																															
	Danger!	Гетр	eratu	res a	bove '	-15°C	c are	too w	arm! \	Write	any c	out-of	-rang	e tem	ıps ar	nd roc	om ter	mp or	n the	lines	belov	v and	call 7	ΓIP im	medi	ately	! 1-8	00-40	4-300)6		
S	-15°C																															
ture	-16°C																															
ھ	-17°C																															
mpe	-18°C																															
tel	-19°C																															
əlc	-20°C																															
ptable	-21°C																															
ce	-22°C																															
ac	-50°C to -23°C																							_								
ction	Out-of-range temps (above -15°C or below -50°C)																															
ac	Room Temperature																									_						

Please complete all information for each excursion outside the proper temperature range; above -15°C or below -50°C.

Day of Month	Time	Temperature	Actions Taken	Signature